

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2014/11/10
Person Assigned	Ulrich Busch (ulrich.busch@varian.com)
Submitter Name	Christof Schadt (christof.schadt@brainlab.com)
Submission Date	2014/03/20

Correction Number	CP-1398
Log Summary: Add FOR Module to RTSS IOD	
Name of Standard PS 3 2014b	
<p>Rationale for Correction:</p> <p>By declaring the Frame of Reference Relationship Sequence (3006,00C0) as deprecated with CP 1249, there is no reason that the Frame of Reference Module is not part of the RT Structure Set Information Object Definition.</p> <p>As it also eases the interpretation of this object by high-level DICOM interpreters, the proposal is to add this Module to the IOD.</p>	
<p>Correction Wording:</p> <p><include proposed change below, following guidelines for formatting of changes in supplements></p>	

Adapt A.19.3 RT Structure Set IOD Module Table

Table A.19.3-1—RT STRUCTURE SET IOD MODULES

IE	Module	Reference	Usage
Patient	Patient	C.7.1.1	M
	Clinical Trial Subject	C.7.1.3	U
Study	General Study	C.7.2.1	M
	Patient Study	C.7.2.2	U
	Clinical Trial Study	C.7.2.3	U
Series	RT Series	C.8.8.1	M
	Clinical Trial Series	C.7.3.2	U
<u>Frame of Reference</u>	<u>Frame of Reference</u>	<u>C.7.4.1</u>	<u>U – See Note 2</u>
Equipment	General Equipment	C.7.5.1	M
Structure Set	Structure Set	C.8.8.5	M
	ROI Contour	C.8.8.6	M
	RT ROI Observations	C.8.8.8	M
	Approval	C.8.8.16	U
	SOP Common	C.12.1	M

Note: 1. The Audio Module was previously included in this IOD but has been retired. See PS 3.3 2004.
2. Previously the Frame of Reference Module was not included (see PS 3.3 2014b). Historically, more than one Frame of Reference could be referenced within the Structure Set Module's Referenced Frame of Reference Sequence (3006,0010), in which case the Frame of Reference Module would not have been present.