

## DICOM Correction Proposal

Status	Final Text
Date of Last Update	2016/03/25
Person Assigned	Ulrich Busch ( <a href="mailto:ulrich.busch@varian.com">ulrich.busch@varian.com</a> )
Submitter Name	Ulrich Busch ( <a href="mailto:ulrich.busch@varian.com">ulrich.busch@varian.com</a> )
Submission Date	2015/06/08

Correction Number	CP-1515
Log Summary:	Addition of Representation for RT Cone Beam Acquisition
Name of Standard	PS3.3 2016a
Rationale for Correction:	<p>Dose Reporting for Imaging in Radiotherapy uses TID 10011 CT Radiation Dose as in other modalities. This TID includes TID 10013. CT Irradiation Event Data.</p> <p>In TID 10013 the concept (113820, DCM, "CT Acquisition Type") has to be declared. At this time there is no appropriate code to annotate the cone beam acquisition used at the Radiotherapy devices.</p> <p>This CP adds this code to the Defined CID 10013 used in the template to support annotating ConeBeam Acquisition.</p>
Correction Wording:	

*In Part 3.16, add the following Code in Appendix B. DCMR Context Groups (Normative):*

### CID 10013 CT Acquisition Type

**Type: Extensible**

**Version: 2007103120160314**

**Table CID 10013. CT Acquisition Type**

Coding Scheme Designator	Code Value	Code Meaning	SNOMED-CT Concept ID	UMLS Concept Unique ID
DCM	113804	Sequenced Acquisition		
SRT	P5-08001	Spiral Acquisition	116152004	C0860888
DCM	113805	Constant Angle Acquisition		
DCM	113806	Stationary Acquisition		
DCM	113807	Free Acquisition		
<b><u>SRT</u></b>	<b><u>R-FB8F1</u></b>	<b><u>Cone Beam Acquisition</u></b>	702569007	