

1	Status	Final Text
2	Date of Last Update	2016/05/25
3	Person Assigned	David Clunie
4		mailto:dclunie@dclunie.com
5	Submitter Name	David Clunie
6		mailto:dclunie@dclunie.com
7	Submission Date	2015/09/28

8	Correction Number CP-1547	
9	Log Summary: Add syringe to device CID	
10	Name of Standard	
11	PS3.16 2016b	
12	Rationale for Correction:	
13	Syringes filled with some substance are often used as phantoms.	
14	CID 4051 is the baseline CID for the Device Sequence in the Device Module, which is described in the Quality Control Image attribute	
15	as being a means of describing the phantom.	
16	Correction Wording:	

Amend DICOM PS3.16 as follows:

CID 4051 General Devices

Type: Extensible
Version: 2006102320160525

Table CID 4051. General Devices

Coding Scheme Designator	Code Value	Code Meaning
<i>Include CID 8 Angiographic Interventional Devices</i>		
<i>Include CID 3451 Calibration Objects</i>		
<i>Include CID 4052 "Phantom Devices"</i>		
<u>SRT</u>	<u>A-10150</u>	<u>Syringe</u>

CID 4052 Phantom Devices

Type: Extensible
Version: 20061023

Table CID 4052. Phantom Devices

Coding Scheme Designator	Code Value	Code Meaning
DCM	113681	Phantom
DCM	113682	ACR Accreditation Phantom - CT
DCM	113683	ACR Accreditation Phantom - MR
DCM	113684	ACR Accreditation Phantom - Mammography
DCM	113685	ACR Accreditation Phantom - Stereotactic Breast Biopsy
DCM	113686	ACR Accreditation Phantom - ECT
DCM	113687	ACR Accreditation Phantom - PET
DCM	113688	ACR Accreditation Phantom - ECT/PET
DCM	113689	ACR Accreditation Phantom - PET Faceplate
DCM	113690	IEC Head Dosimetry Phantom
DCM	113691	IEC Body Dosimetry Phantom
DCM	113692	NEMA XR21-2000 Phantom