

### DICOM Correction Item

Correction Number		CP-157 (Final Text)
Log Summary: Correct typos in US Image Module		
Type of Modification	Name of Standard	
Correct typo	PS 3.3 - 1998	
Rationale for Correction		
Typos in the Ultrasound Supplement 5 have been propagated into the 1996 and 1998 standard unnoticed until now:		
Table C.8-18		
The Attribute Description for Bits Allocated is wrong.		
The Attribute Description for Bits Stored is wrong.		
Section C.8.5.6.1.4		
00180063 should be 00181063		
Sections of documents affected		
Section C.8.5.6, Table C.8-18		
Section C.8.5.6.1.4		
Correction Wording:		

#### C.8.5.6 US Image Module

Table C.8-18 specifies the Attributes that describe ultrasound images.

**Table C.8-18**  
**US IMAGE MODULE ATTRIBUTES**

...			
Photometric Interpretation	(0028,0004)	1	Specifies the intended interpretation of the pixel data. See C.8.5.6.1.2 for specialization.
Bits Allocated	(0028,0100)	1	<del>Specifies the intended interpretation of the pixel data.</del> Number of bits allocated for each pixel sample. See C.8.5.6.1.13 for specialization.
Bits Stored	(0028,0101)	1	<del>Number of bits allocated for each pixel sample.</del> Number of bits stored for each pixel sample. See C.8.5.6.1.14 for specialization.
High Bit	(0028,0102)	1	Most significant bit for pixel sample data. See C.8.5.6.1.15 for specialization.
...			

...

**C.8.5.6.1.4 Frame Increment Pointer**

For US Multi-frame images, the Attribute Frame Increment Pointer (0028,0009) of the Multi-frame Module (see Section C.7.6.6) is specified by the following Defined Terms:

~~**00180063 = sequencing by Frame Time (0018,1063)**~~

**00181063 = sequencing by Frame Time (0018,1063)**

00181065 = sequencing by Frame Time Vector (0018,1065)