

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2017/06/05
Person Assigned	Kevin O'Donnell
Submitter Name	Kevin O'Donnell
Submission Date	2016/11/08

Correction Number	CP-1670
Log Summary: Add number of items for Responsible Group Code Sequence	
Name of Standard PS3.3 2018b	
Rationale for Correction: Responsible Group Code Sequence (0008,0220) is missing number of items permitted in sequence in Table C.34.2-1. Protocol Context Module Attributes in PS3.3.	
Correction Wording:	

Modify Table C.34.2-1 as shown:

Table C.34.2-1. Protocol Context Module Attributes

Attribute Name	Tag	Type	Attribute Description
Custodial Organization Sequence	(0040,A07C)	3	Custodial organization for this Protocol instance. Represents the organization that is currently in charge of maintaining this protocol instance. <i>Note</i> <i>This may or may not be identical to the Institution identified in the Equipment Module. This may or may not be the Institution that originally created this instance.</i> Only a single Item is permitted in this Sequence.
>Institution Name	(0008,0080)	2	Name of Custodial Institution or Organization.
>Institution Code Sequence	(0008,0082)	2	Coded identifier of Custodial Institution or Organization. Zero or one Item shall be included in this Sequence.
>>Include Table 8.8-1 "Code Sequence Macro Attributes"		No Baseline CID is defined.	
Responsible Group	(0008,0220)	2	The department, unit or service that is responsible

Attribute Name	Tag	Type	Attribute Description
Code Sequence			for the management of this Protocol. See Section C.34.2.3 . <u>Zero or more Items shall be included in this Sequence.</u>
>Include Table 8.8-1 “Code Sequence Macro Attributes”			Baseline CID 7030 “Institutional Departments, Units and Services” .
...			