

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2017/09/14
Person Assigned	Rob Horn
Submitter Name	Rob Horn
Submission Date	2017/01/23

Correction Number	CP-1686
Log Summary:	Add UDI code for use in templates
Name of Standard	PS3.16 2017c
Rationale for Correction:	<p>The Unique Device Identifier (UDI) of a device or drug may be needed as contents within a structured report, e.g., as used in Sup 164 templates.</p> <p>This adds a DICOM code to indicate that a field is a UDI.</p>
Correction Wording:	

**Add to PS 3.16 Annex D**

Code Value	Code Meaning	Definition	Notes
128739	UDI	<p>The entire Human Readable Form of the Unique Device Identifier as defined by the Issuing Agency.</p> <p>See PS3.3 Section 10.29.1.</p>	