

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2017/09/17
Person Assigned	Wim Corbijn
Submitter Name	Wim Corbijn
Submission Date	2017/01/25

Correction Number	CP-1696
Log Summary:	Clarify condition for Repetition Time (0018,0080)
Name of Standard	PS3.3 2017c
Rationale for Correction:	<p>Current description of the condition is negatively phrased which makes it more difficult to understand. It also has no statement if it can be present otherwise or not.</p> <p>Make the required statement positive and add the statement that the tag “may be present otherwise”.</p>
Correction Wording:	

*Item #1: Update Table C.8-4 Modify Required clause for Repetition Time (0018,0080)*

Note

**Table C.8-4. MR Image Module Attributes**

Attribute Name	Tag	Type	Attribute Description
....			
Repetition Time	(0018,0080)	2C	<p>The period of time in msec between the beginning of a pulse sequence and the beginning of the succeeding (essentially identical) pulse sequence. <del>Required except when Scanning Sequence (0018,0020) is EP and Sequence Variant (0018,0021) is not SK.</del></p> <p><b>Required if Sequence Variant (0018,0021) is SK or if Scanning Sequence (0018,0020) is not EP.</b></p> <p><b><u>May be present otherwise.</u></b></p>
.....			