

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2018/06/05
Person Assigned	Wim Corbijn
Submitter Name	Nikhilesh Sonar ( <a href="mailto:nikhilesh.sonar@philips.com">nikhilesh.sonar@philips.com</a> )
Submission Date	2017/11/03

Correction Number	CP-1751
Log Summary: MR Spectroscopy Functional Group Macros definition contains an improper reference of a Macro	
Name of Standard	PS3.3 2018b
Rationale For Correction: MR Spectroscopy Functional Group Macros definition (Section A.36.3.4) contains a reference of the MR Spectroscopy Frame Type (a Macro), but the reference (C.8.14.5.1) points to a descriptive section and not a table of attributes.  Possible correction of the macro reference might be the MR Spectroscopy Frame Type Macro (Section C.8.14.3.1), of the same name. This Macro is currently not referred by any IOD in the Standard.	
Correction Wording:	

*Item #1: Update Table A.36-4. MR Spectroscopy Functional Group Macros*

**Table A.36-4. MR Spectroscopy Functional Group Macros**

Functional Group Macro	Section	Usage
...		
Respiratory Synchronization	C.7.6.16.2.17	C - Required if Respiratory Motion Compensation Technique (0018,9170) equals other than NONE, REALTIME or BREATH_HOLD and if Image Type (0008,0008) Value 1 is ORIGINAL or MIXED. May be present otherwise.
MR Spectroscopy Frame Type	<del>C.8.14.5.1</del> <u>C.8.14.3.1</u>	M
MR Timing and Related Parameters	C.8.13.5.2	C - Required if Image Type (0008,0008) Value 1 is ORIGINAL or MIXED. May be present otherwise.
...		