

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2018/06/05
Person Assigned	Wim Corbijn
Submitter Name	Nikhilesh Sonar (nikhilesh.sonar@philips.com)
Submission Date	2017/11/03

Correction Number	CP-1753
Log Summary: Tractography Results IOD definition incorrectly references a retired Module	
Name of Standard	PS3.3 2018b
<p>Rationale For Correction:</p> <p>Tractography Results IOD definition (Section A.78.3) contains a reference of the Specimen Module, but the reference (C.7.1.2) points to the Specimen Identification Module, and then redirects to the Specimen Module (Section C.7.6.22) – citing a reason that the Specimen Identification Module was retired in the year 2008.</p> <p>If the original intent was to include the Specimen Module in the IOD, then the indirection to the Specimen Identification Module would seem incorrect.</p>	
Correction Wording:	

Item #1: Update Table A.1-3 to add a Specimen row for the 'Trc Res' column

Table A.1-3. Composite Information Object Modules Overview - More Non-Images

IODs Modules	Trc Res
Patient	M
Clinical Trial Subject	U
General Study	M
Patient Study	U
Clinical Trial Study	U

IODs Modules	Trc Res
General Series	M
Clinical Trial Series	U
...	...
Tractography Results Series	M
Frame of Reference	M
General Equipment	M
Enhanced General Equipment	M
<u>Specimen</u>	<u>U</u>
Palette Color LUT	
...	
SOP Common	M

Item #2: Update A.78-1. Tractography Results IOD Modules

Table A.78-1. Tractography Results IOD Modules

IE	Module	Reference	Usage
Patient	Patient	C.7.1.1	M
	<u>Specimen</u>	<u>C.7.1.2</u>	<u>U</u>

IE	Module	Reference	Usage
	Clinical Trial Subject	C.7.1.3	U
Study	General Study	C.7.2.1	M
	Patient Study	C.7.2.2	U
	Clinical Trial Study	C.7.2.3	U
Series	General Series	C.7.3.1	M
	Tractography Results Series	C.8.33.1	M
	Clinical Trial Series	C.7.3.2	U
Frame of Reference	Frame of Reference	C.7.4.1	M
Equipment	General Equipment	C.7.5.1	M
	Enhanced General Equipment	C.7.5.2	M
Tractography Results	Tractography Results	C.8.33.2	M
	<u>Specimen</u>	<u>C.7.6.22</u>	<u>U</u>
	Common Instance Reference	C.12.2	M
	SOP Common	C.12.1	M