

DICOM Correction Item

| | |
|--|-------------------------------|
| Log Summary: Correct Results Interpretation Relationship Module | |
| Type of Modification Correction | |
| Name of Document PS 3.3 | Version Number 1999 |
| Rationale for change In the 1996 version of the standard, the relationship in the DICOM Information Model between Results IOD and Interpretation IOD is 0-1 (see PS3.3 Figure 7-2). In the Results Relationship Module (PS3.3 Table C.5-1), the Referenced Interpretation Sequence seems to imply multiple items are allowed, but this conflicts Figure 7-2. In the 1998 version of the standard, the Referenced Interpretation Sequence is clarified to specify zero or more items are allowed. The figure, now 7-2a, is not changed. | |
| Sections of document affected PS3.3 Figure 7-2a | |
| Changes Changes Figure 7-2a to specify the relationship between Results IOD and Interpretation IOD to 0-n | |

