

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2023/01/21
Person Assigned	Harry Solomon
Submitter Name	Harry Solomon
Submission Date	2022/06/09

Correction Number	CP-2239
Log Summary:	Institution Name is not Issuer of ID
Name of Standard	PS3.3 2022e
Rationale for Correction:	<p>The Equipment IE includes the Attribute Institution Name (0008,0080) “Institution where the equipment that produced the Composite Instances is located.” It should be clarified that this Attribute in the Equipment IE should not be construed as the Issuer of Patient ID or Accession Number.</p> <p>The Visit IE, used in the Modality Worklist Information Model, similarly includes Institution Name defined as “Institution where the equipment is located.” While this is not properly modeled as part of a “Visit” (which in the broader sense does not necessarily involve [imaging] equipment), it is clear that even in that IE it is not the Issuer of Patient ID or Accession Number, which are part of the Patient IE and Imaging Service Request IE, respectively. See, e.g., <a href="#">PS3.4 Figure K.6-1. Modality Worklist Information Model E-R Diagram</a>, where the Visit IE has no E-R relationship to the Patient and Imaging Service Request IEs.</p>
Correction Wording:	

**Table C.7-8. General Equipment Module Attributes**

Attribute Name	Tag	Type	Attribute Description
Manufacturer	(0008,0070)	2	Manufacturer of the equipment that produced the Composite Instances.
Institution Name	(0008,0080)	3	<p>Institution where the equipment that produced the Composite Instances is located.</p> <p><b>Note</b></p> <p><b><u>This Attribute represents the organizational context only for the Equipment IE, and should not be construed to be a substitute for Issuer of Patient ID (0010,0021) or Issuer of Accession Number (0008,0051).</u></b></p>
...			