

### DICOM Correction Item

Correction Number		CP-242
Log Summary: Duplicate NM retired/new elements in data dictionary		
Type of Modification	Name of Standard	
Correction	PS 3.6 2000	
Rationale for Correction		
Some data elements were renamed and had their value multiplicity changed in PS 3.6 when the revised NM object was folded into the standard, but are still present (duplicated) and flagged as "retired". This leads to problems when automatically generating code from the tables. It is proposed the retired duplicates be deleted.		
Sections of documents affected		
PS 3.6		
Correction Wording:		

<del>(0018,1070)</del> <b>Radionuclide Route</b>	<b>LO</b>	<b>1-n</b>	<b>RET</b>
(0018,1070) Radiopharmaceutical Route	LO	1	
<del>(0018,1071)</del> <b>Radionuclide Volume</b>	<b>DS</b>	<b>1-n</b>	<b>RET</b>
(0018,1071) Radiopharmaceutical Volume	DS	1	
<del>(0018,1072)</del> <b>Radionuclide Start Time</b>	<b>TM</b>	<b>1-n</b>	<b>RET</b>
(0018,1072) Radiopharmaceutical Start Time	TM	1	
<del>(0018,1073)</del> <b>Radionuclide Stop Time</b>	<b>TM</b>	<b>1-n</b>	<b>RET</b>
(0018,1073) Radiopharmaceutical Stop Time	TM	1	
<del>(0018,1074)</del> <b>Radionuclide Total Dose</b>	<b>DS</b>	<b>1-n</b>	<b>RET</b>
(0018,1074) Radionuclide Total Dose	DS	1	