DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2026/03/21
Person Assigned	Christof Schadt
Submitter Name	Kari Jyrkkälä <kari.jyrkkala@varian.com>
Submission Date	2025/04/22

Correction Number CP-2560
Log Summary: General Anatomy Optional Macro in RT Structure Set IOD
Name of Standard PS3.3 2026a
Rationale for Correction: General Anatomy Optional Macro in General Image Module can be used to identify the anatomic region of the acquired images. Treatment Site Code Sequence (3010,0078) in RT General Plan Module can be used for similar purposes. However, in RT Structure Set IOD there is no way of indicating the anatomic region, thus General Anatomy Optional Macro is introduced to Structure Set Module.
Correction Wording:

Update PS3.3, C.8.8.5 Structure Set Module

Table C.8-41. Structure Set Module Attributes

Attribute Name	Tag	Type	Attribute Description
Structure Set Label	(3006,0002)	1	User-defined label for Structure Set.
Structure Set Name	(3006,0004)	3	User-defined name for Structure Set.
Structure Set Description	(3006,0006)	3	User-defined description for Structure Set.
Instance Number	(0020,0013)	3	A number that identifies this object Instance.
Structure Set Date	(3006,0008)	2	Date at which the content of the Structure Set was last modified.
Structure Set Time	(3006,0009)	2	Time at which the content of the Structure Set was last modified.
Include Table 10-7 "General Anatomy Optional Macro Attributes"			***Anatomic Region Sequence BCID 4 "Anatomic Region".***
Referenced Frame of Reference Sequence	(3006,0010)	3	Sequence describing Frames of Reference in which the ROIs are defined. One or more Items are permitted in this Sequence. See Section C.8.8.5.1.
>Frame of Reference UID	(0020,0052)	1	Uniquely identifies Frame of Reference within Structure Set.