

DICOM Correction Proposal Form

Correction Number CP-305	
Log Summary: Correct error in PS 3.16, TID 4006	
Type of Modification Correction	Name of Standard PS 3.16
Rationale for Correction There is an error in PS 3.16, TID 4006 Mammography CAD Single Image Finding. Row 17 was mistakenly assigned a Relationship with Parent of "R-HAS PROP". The correct relationship is "HAS PROPERTIES".	
Sections of documents affected PS 3.16, Annex A	
Correction Wording: Update PS 3.16, Annex A, TID 4006, row 17:	

**TID 4006
MAMMOGRAPHY CAD SINGLE IMAGE FINDING**

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1			CODE	(111059, DCM, "Single Image Finding")	1	M		DCID (6014) "Mammography Single Image Finding"
...								
17	>	R-HAS PROPERTIES	SCoord	(111030, DCM, "Image Region")	1-n	MC	Shall be present only if value of parent is (111101, DCM, "Image quality") and row 16 is not present	
18	>>	R-SELECTED FROM	IMAGE		1	M		All the (111030, DCM, "Image Region") content items in a single invocation of this template shall reference the same IMAGE content item in the (111028, DCM, "Image Library")
...								