

DICOM Correction Proposal Form

Correction Number CP-321	
Log Summary: Key object selection for printing	
Type of Modification Correction	Name of Standard PS 3.16 2001
Rationale for Correction Users have expressed the desire to flag images using key object selections for subsequent printing. How the objects are actually scheduled to be printed (manually or automatically) is specific to the application outside the scope of the standard.	
Sections of documents affected PS 3.16	
Correction Wording:	

Correct the following code to CID 7010 in PS 3.16:

Context ID 7010
Key Object Selection Document Title
(Most Restrictive Use: Defined)

Coding Scheme Designator (0008,0102)	Code Value (0008,0100)	Code Meaning (0008,0104)
DCM	113000	Of Interest
DCM	113001	Rejected for Quality Reasons
DCM	113002	For Referring Provider
DCM	113003	For Surgery
DCM	113004	For Teaching
DCM	113005	For Conference
DCM	113006	For Therapy
DCM	113007	For Patient
DCM	113008	For Peer Review
DCM	113009	For Research
DCM	113010	Quality Issue
DCM	113013	Best In Set
DCM	113018	For Printing

Correct the following definition to PS 3.16 Annex D:

113011	Document Title Modifier	Document Title Modifier
113012	Key Object Description	Key Object Description
113013	Best In Set	
113018	For Printing	For Printing
113026	Double exposure	Double exposure
121001	Quotation Mode	