

DICOM Correction Proposal Form

Correction Number		CP-364	
Log Summary: MR Spatial Saturation Sequence incorrect number of items			
Type of Modification		Name of Standard	
Clarification		PS 3.3 2003	
Rationale for Correction Multiple spatial saturation slabs (e.g. in different directions or at multiple ends of a volume) may be applied in practice but the sequence in which they are described allows only zero or one item, whereas more than one is necessary.			
Sections of documents affected PS 3.3 C.8.13.5.11			
Correction Wording:			

Table C.8.13-25
MR SPATIAL SATURATION MACRO ATTRIBUTES

Attribute Name	Tag	Type	Attribute Description
MR Spatial Saturation Sequence	(0018,9107)	2	A sequence that provides the position of spatial saturation bands deposited as part of the pulse sequence. Zero or one more Items may be included in this sequence.