

DICOM Correction Proposal Form

Correction Number		CP-370
Log Summary: Wrong VR for LUT Label in Enhanced MR		
Type of Modification	Name of Standard	
Clarification	PS 3.3 2003	
Rationale for Correction		
The VR of LUT Label added for the Real World Value mapping Sequence is SS , a binary signed short, which is clearly not what was intended. It is changed to the more appropriate SH, short string.		
Sections of documents affected		
PS 3.6		
Correction Wording:		

(0040,9210) LUT Label

SS SH 1