

1. 2 **DICOM Correction Item**

Correction Number		CP-405
Log Summary: Remove the restriction of VOI LUT from the NM IOD		
Type of Modification	Name of Standard	
Remove VOI LUT module from NM IOD	PS 3.3 2003	
Rationale for Correction		
For the consistency with the NM Image IOD Entity-Relationship Model, the VOI LUT need NOT been removed from the NM Image IOD Modules in section A.5.2		
Sections of documents affected		
PS 3.3 Sections: A.5.2		
Correction Wording:		

1. 4

1. 6 ***Remove the restriction of VOI LUT module from NOT being included in the NM IOD in A.5.2***

1. 8 **A.5.2 NM Image IOD Entity-Relationship Model**

1. 10 The E-R Model in Section A.1.2 depicts those components of the DICOM Information Model which directly reference the NM Image IOD. The Modality LUT IE ~~and VOI LUT IE are~~ **is** not a components of the NM Image IOD.

1. 12