

### DICOM Correction Item

Correction Number		CP-445
Log Summary: Allows X-Ray Table Module included when only angulation done		
Type of Modification	Name of Standard	
Change	PS 3.3 2003	
Rationale for Correction		
<p>In the XA IOD it is not allowed to specify a table angulation (attribute Table Angle (0018,1138) when not table motion is applied.</p> <p>With this change proposal the module is allowed for this cases.</p>		
Sections of documents affected		
PS 3.3 Section A14.3		
Correction Wording:		

#### A.14 X-RAY ANGIOGRAPHIC IMAGE INFORMATION OBJECT DEFINITION

##### A.14.3 XA Image IOD Module Table

Table A.14.-1  
 X-RAY ANGIOGRAPHIC IMAGE IOD MODULES

IE	Module	Reference	Usage
	...		
	X-Ray Table	C.8.7.4	C - Required if Image is created with table motion, <b><u>may be present otherwise.</u></b>
	...		