

DICOM Correction Item

Correction Number CP-451	
Log Summary: Add Key Object Selection Document Title for Report Attachment	
Type of Modification Modification	Name of Standard PS 3.16
Rationale for Correction: A major major increment in radiology reporting functionality can be achieved by a simple, automated process that can attach key images to a clinical report. The problem today is that transcription of reports comes through a system path disjoint from the image review path. A Key Object Selection Document, created at the image review station and identifying images to be appended to a report, can be used to pass image references to the report management system; those images can then be appended to the report by an automated process after transcription is complete.	
Sections of documents affected PS 3.16 Annex B and Annex D	
Correction Wording:	

Annex B

CID 7010 Key Object Selection Document Title

**Context ID 7010
 Key Object Selection Document Title**

Type: Extensible Version: 2002090420040920

Coding Scheme Designator (0008,0102)	Code Value (0008,0100)	Code Meaning (0008,0104)
...		
<u>DCM</u>	<u>113020</u>	<u>For Report Attachment</u>

Annex D

DICOM Code Definitions (Coding Scheme Designator "DCM" Coding Scheme Version "01")

Code Value	Code Meaning	Definition	Notes
...			
<u>113020</u>	<u>For Report Attachment</u>	<u>Selection of information objects for attachment to the clinical report of the Current Requested Procedure</u>	