

DICOM Correction Item

Correction Number CP-458	
Log Summary: Correct condition in TID 1404	
Type of Modification Modification	Name of Standard PS 3.16 2003
Rationale for Correction: The requirement type for row 6 in TID 1404 should be "UC" rather than "U" because of the given condition.	
Sections of documents affected PS 3.16 Annex B	
Correction Wording:	

TID 1404
NUMERIC MEASUREMENT
Type: Extensible

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
6	>	INFERRED FROM	IMAGE	BCID (7003) Diagnostic Imaging Report Purposes of Reference	1-n	<u>UC</u>	XOR Row 2,5	