

DICOM Correction Item

Correction Number	CP-477
Log Summary:	Add PPS Discontinued Reason – Equipment Change
Type of Modification	Name of Standard
Addition	PS 3.16-2004
Rationale for Correction: The IHE Cardiology Technical Committee has identified a use case for discontinuing a Performed Procedure Step when different equipment is desired to be used, e.g., because of additional features or capabilities, not necessarily because of an equipment failure. This proposal adds a new PPS Discontinued Reason Code.	
Sections of documents affected PS 3.16 Annex B and D	
Correction Wording:	
<i>Item #1: Modify Part 16 Annex B CID 9300:</i>	

CID 9300 Procedure Discontinuation Reasons

**Context ID 9300
 Procedure Discontinuation Reasons**

Type: Extensible Version: 2002090420050110

Coding Scheme Designator (0008,0102)	Code Value (0008,0100)	Code Meaning (0008,0104)
...		
DCM	110516	Equipment change

<i>Item #2: Modify Part 16 Annex D:</i>

DICOM Code Definitions (Coding Scheme Designator “DCM” Coding Scheme Version “01”)

Code Value	Code Meaning	Definition	Notes
...			
110516	Equipment change	Procedure step is discontinued to change to other equipment or modality	
...			