

DICOM Correction Item

Correction Number CP-478	
Log Summary: Move term from CID 6054 to CID 6016	
Type of Modification	Name of Standard
Correction	PS 3.16 - 2004
Rationale for Correction: A composite feature term is needed in CID 6016 in PS 3.16, Annex B, that indicates a mass that contains calcifications, for use in Mammography CAD SR. An existing term is being moved from CID 6054 to CID 6016, so that it is accessible to both Breast Imaging Report and Mammography CAD SR.	
Sections of documents affected PS 3.16, Annex B	
Correction Wording: Remove "Mass with calcifications" term from CID 6054 in PS 3.16, Annex B. Add "Mass with calcifications" term to CID 6016 in PS 3.16, Annex B.	

CID 6016 Mammography Composite Feature

CONTEXT ID 6016

Mammography Composite Feature

Type: Extensible Version: 20050110

Note: In future extensions, Mammography Composite Feature terms that are not derived from BI-RADS® should be added to this context group.

Coding Scheme Designator (0008,0102)	Coding Scheme Version (0008,0103)	Code Value (0008,0100)	Code Meaning (0008,0104)
DCM		111459	Mass with calcifications
<i>Include CID 6014</i>			
<i>Include CID 6017</i>			

CID 6054 BREAST IMAGING FINDINGS

CONTEXT ID 6054

Breast Imaging Findings

Type: Extensible Version: 20050110

Coding Scheme Designator (0008,0102)	Coding Scheme Version (0008,0103)	Code Value (0008,0100)	Code Meaning (0008,0104)
SRT		F-8A084	Breast normal
SRT		F-8A057	Calcification of breast
DCM		111459	Mass with calcifications
SRT		A-04010	Implant
<i>Include CID 6016</i>			
<i>Include CID 6057</i>			
<i>Include CID 6064</i>			