

CP-609

Date: 2006/03/27

Replace inconsistent retired indicator for elements (0018,9195) and (0018,9196)

Status: Final Text

DICOM Correction Item

Correction Number		CP-609
Log Summary: Replace inconsistency retired indicator for elements (0018,9195) and (0018,9196)		
Type of Modification	Name of Standard	
Correction	PS 3.6 2004	
Rationale for Correction.		
Elements (0018,9195) and (0018,9196) are marked as retired by "Retired" instead of "RET".		
Sections of documents affected		
PS 3.6 Section 6		
Correction Wording:		

<i>In PS3.6, Section 6, replace "Retired" by "RET":</i>

Tag	Name	VR	VM	
(0018,9195)	Chemical Shifts Minimum Integration Limit in Hz	FD	1	Retired <u>RET</u>
(0018,9196)	Chemical Shifts Maximum Integration Limit in Hz	FD	1	Retired <u>RET</u>
(0028,9099)	Largest Monochrome Pixel Value	US	1	Retired <u>RET</u>