

## DICOM Correction Item

Correction Number	CP-629
Log Summary: Correct cardiology units of measurement	
Type of Modification	Name of Standard
Correction	PS 3.16 2006
Rationale for Correction	
The prefix for kilohertz is listed incorrectly in CID 3082.	
Sections of documents affected	
PS 3.16 CID 3082	
Correction Wording:	

### **CID 3082      Cardiology Units of Measurement**

#### **Context ID 3082 Cardiology Units of Measurement**

**Type: Extensible Version: ~~20020904~~20060613**

Coding Scheme	Coding Scheme Version	Code Value	Code Meaning
UCUM	1.4	J	Joule
UCUM	1.4	<del>K</del> Hz	kiloHerz
UCUM	1.4	kOhm	kiloOhm