

DICOM Correction Item

Correction Number		CP-645
Log Summary: Specify number of items for Device and Intervention Sequences		
Type of Modification	Name of Standard	
Correction	PS 3.3 2006	
Rationale for Correction		
The number of items is not specified for Device Sequence in the Device Module, nor for the Intervention Sequence in the Intervention Module, and both should be one or more.		
The number of items for Intervention Drug Sequence in the Intervention Module is not specified, and should be exactly one.		
Sections of documents affected		
PS 3.3 C.7.6.12, C.7.6.13		
Correction Wording:		

C.7.6.12 Device Module

...

Table C.7-18
DEVICE MODULE ATTRIBUTES

Attribute Name	Tag	Type	Attribute Description
Device Sequence	(0050,0010)	3	Introduces sequence of items describing devices used that may be visible in the image. <u>One or more Items may be included in this Sequence.</u>
>Include 'Code Sequence Macro' Table 8.8-1		Baseline Context ID is 8.	
...

C.7.6.13 Intervention Module

...

Table C.7-19
INTERVENTION MODULE ATTRIBUTES

Attribute Name	Tag	Type	Attribute Description
Intervention Sequence	(0018,0036)	3	Introduces sequence of items describing interventional therapies or procedures. <u>One or more Items may be included in this Sequence.</u>
>Include 'Code Sequence Macro' Table 8.8-1		Baseline Context ID is 9.	
>Intervention Status	(0018,0038)	2	Temporal relation of SOP Instance to

			intervention Enumerated Values: PRE INTERMEDIATE POST NONE
>Intervention Drug Sequence	(0018,0029)	3	Sequence that identifies the interventional drug. <u>Only one item shall be included in this Sequence.</u>
>>Include 'Code Sequence Macro' Table 8.8-1		Baseline Context ID is 10.	
>Intervention Drug Start Time	(0018,0035)	3	Time of administration of the interventional drug.
>Intervention Drug Stop Time	(0018,0027)	3	Time of completion of administration of the intervention drug.
> Administration Route Code Sequence	(0054,0302)	3	Sequence that identifies the Administration Route. This sequence shall contain exactly one item.
>>Include 'Code Sequence Macro' Table 8.8-1		Baseline Context ID is 11.	
>Intervention Description	(0018,003A)	3	Further description in free form text describing the therapy or other intervention.