

DICOM Correction Item

Correction Number		CP-682
Log Summary: Align description annotation attribute enh. XA/XRF		
Type of Modification	Name of Standard	
Correction	PS 3.3 - 2007	
Rationale for Correction		
<p>CP 627 introduced the Burned In Annotation attribute for Enhanced MR and CT. The description of this attribute is different with the one of the Enhanced XA/XRF IOD.</p> <p>This CP aligns the description.</p>		
Sections of documents affected		
PS 3.3 Section C.8.19.2		
Correction Wording:		

C.8.19.2 Enhanced XA/XRF Image Module

...

Table C.8.19.2-1
Enhanced XA/XRF Image Module Table

Attribute Name	Tag	Type	Attribute Description
...			
Burned In Annotation	(0028,0301)	1	<p>Indicates that the image shall not contain burned in annotations<u>whether or not the image contains sufficient burned in annotation to identify the patient and date the image was acquired.</u></p> <p>Enumerated Values: NO</p>
...			