

## DICOM Correction Item

|   |                  |        |
|---|------------------|--------|
| Correction Number   |                  | CP 754 |
| Log Summary: Add Display Shutter Module to CR IOD                                     |                  |        |
| Type of Modification  | Name of Standard |        |
| Addition  | PS 3.3 2007      |        |
| Rationale for Correction  |                  |        |
| The Display Shutter is useful for CR applications, and often added by vendors anyway. |                  |        |
| Sections of documents affected  |                  |        |
| PS 3.3 A.2.3  |                  |        |
| Correction Wording:   |                  |        |

*Add the Display Shutter Module to overview table and CR Image IOD Module Table:*

### A.2.3 CR Image IOD Module Table

**Table A.2-1  
 CR IMAGE IOD MODULES**

| IE    | Module                        | Reference              | Usage   |
|-------|-------------------------------|------------------------|---|
| ...   | ...                           | ...                    | ...   |
| Image | General Image                 | C.7.6.1                | M   |
|       | Image Pixel                   | C.7.6.3                | M   |
|       | Contrast/bolus                | C.7.6.4                | C - Required if contrast media was used in this image |
|       | <b><u>Display Shutter</u></b> | <b><u>C.7.6.11</u></b> | <b><u>U</u></b>                                       |
|       | Device                        | C.7.6.12               | U   |
| ...   | ...                           | ...                    | ...   |