

DICOM Correction Item

Correction Number		CP-808
Log Summary: Change Type Acquisition Duration for Enhanced CT		
Type of Modification	Name of Standard	
Modification	PS 3.3 – 2007	
Rationale for Correction		
<p>In the Enhanced CT the Acquisition Duration (0018,9073) is define as Type 2C while in Enhanced MR and PET it is a Type 1C attribute.</p> <p>Proposed is to harmonize the Attribute Type. No implementations of Enhanced CT are existing, this change do not break implementations.</p>		
Sections of documents affected		
PS 3.3 C.8.15.2		
Correction Wording:		

C.8.15.2 Enhanced CT Image Module

This section describes the Enhanced CT Image Module. Table C.8-114 specifies the attributes of the Enhanced CT Image Module.

**Table C.8-114
 ENHANCED CT IMAGE MODULE ATTRIBUTES**

Attribute Name	Tag	Type	Attribute Description
...			
Acquisition Duration	(0018,9073)	21C	<p>The time in seconds needed to complete the acquisition of data. See C.7.6.16.2.2.1 for further explanation.</p> <p>Required if Image Type (0008,0008) Value 1 is ORIGINAL or MIXED, may be present otherwise.</p>
...			