

### DICOM Correction Item

Correction Number		CP-856
Log Summary: Missing General Series Module for Deformable Spatial Registration IOD.		
Type of Modification	Name of Standard	
Correction	PS 3.3 – 2008	
Rationale for Correction		
<p>During editing for Final Text of Supplement 112 (Deformable Spatial Registration) the General Series module was removed from the required Deformable Spatial Registration IOD Modules. The General Series Module should have been Mandatory.</p> <p>The General Series Module was also left out of Table A.1-2 for the Spatial Registration IOD.</p>		
Sections of documents affected		
PS 3.3 A.1.4 and A.39.2.1.2		
Correction Wording:		

**Item: Edit the following columns in table A.1-2.**

**Table A.1-2  
 COMPOSITE INFORMATION OBJECT MODULES OVERVIEW - NON-IMAGES**

IODs Modules	REG	DREG
Patient	M	M
Clinical Trial Subject	U	U
General Study	M	M
Patient Study	U	U
Clinical Trial Study	U	U
General Series	M	<b><u>M</u></b>
Clinical Trial Series	U	U
Spatial Registration Series	M	M
Frame Of Reference	M	M
General Equipment	M	M
Enhanced Equipment		M
Spatial Registration	M	
Deformable Spatial Registration		M
Common Instance Reference	M	M
SOP Common	M	M

*Item: Edit Table A.39.2-1.*

**A.39.2.1.2 Deformable Spatial Registration IOD Module Table**

**Table A.39.2-1 DEFORMABLE SPATIAL REGISTRATION IOD MODULES**

<b>IE</b>	<b>Module</b>	<b>Reference</b>	<b>Usage</b>
Patient	Patient	C.7.1.1	M
	Clinical Trial Subject	C.7.1.3	U
Study	General Study	C.7.2.1	M
	Patient Study	C.7.2.2	U
	Clinical Trial Study	C.7.2.3	U
Series	<b><u>General Series</u></b>	<b><u>C.7.3.1</u></b>	<b><u>M</u></b>
	Clinical Trial Series	C.7.3.2	U
	Spatial Registration Series	C.20.1	M
Frame of Reference	Frame of Reference	C.7.4.1	M
Equipment	General Equipment	C.7.5.1	M
	Enhanced Equipment	C.7.5.2	M
Deformable Registration	Deformable Spatial Registration	C.20.3	M
	Common Instance Reference	C.12.2	M
	SOP Common	C.12.1	M