

## DICOM Correction Item

Correction Number		CP-861
Log Summary: Constrain Ophthalmic Tomography Image Type		
Type of Modification	Name of Standard	
Addition	PS 3.3 2008	
Rationale for Correction		
<p>The Ophthalmic Tomography IOD does not include the General Image Module, and the Image Type Attribute is specified in the Ophthalmic Tomography Image Module as Type 1 but with no constraints on value multiplicity or enumerated values. This is inconsistent with all other image types.</p> <p>Constrain the values to the normal values.</p>		
Sections of documents affected		
PS 3.3 C.8.17.7		
Correction Wording:		

*Amend PS 3.3 C.8.17.7:*

**C.8.17.7      Ophthalmic Tomography Image Module**

...

**Table C.8.17.7-1  
 OPHTHALMIC TOMOGRAPHY IMAGE MODULE ATTRIBUTES**

Attribute Name	Tag	Type	Attribute Description
Image Type	(0008,0008)	1	Image identification characteristics. <b><u>See C.7.6.1.1.2 for Enumerated Values and Defined Terms and further explanation.</u></b>
...	...	...	...