

DICOM Correction Item

Correction Number CP-915	
Log Summary: Generalize Beam Delivery Request Object for RT Ion Plans	
Type of Modification Correction	Name of Standard PS 3.3 (Supplement 74)
Rationale for Correction: This object can be and should be available for Proton Delivery as well. Therefore all references to RT Plan should be extended to include RT Ion Plan.	
Sections of documents affected Supplement 74 - Frozen Draft for Trial Implementation, 2007/10/08	
Correction Wording:	

In Supplement 74, C.8.8.aa RT Beams Delivery Instruction Module, Table C.8.8.aa-1 (RT BEAMS DELIVERY INSTRUCTION MODULE ATTRIBUTES), change the following wording

Attribute Name	Tag	Type	Attribute Description
>Referenced RT Plan Sequence	(300C,0002)	1	Reference to a single RT Plan or RT Ion Plan SOP Instance (whose UID is also supplied in the Input Information Sequence – see PS3.4) containing all the Beams and the Fraction Group referenced in this SOP Instance.
>Referenced Beam Number	(300C,0006)	1	Uniquely identifies the Beam which is specified by Beam Number (300A,00C0) within Beam Sequence (300A,00B0) in RT Beams Module of referenced RT Plan or RT Ion Plan.