

## DICOM Correction Item

Correction Number	CP- 976
Log Summary:	Correct Sup 122 Template references
Type of Modification	Name of Standard
Modification	PS 3.3-2008 + Supplement 122
Rationale for Correction:	
Typos.	
Sections of documents affected	
PS 3.3 Annex C	
Correction Wording:	

**Table C.7.6.2x-2  
 SPECIMEN MACRO**

Attribute Name	Tag	Type	Attribute Description
...			
Specimen Description Sequence	(0040,0560)	1	Sequence of identifiers and detailed description of the specimen(s) being imaged. One or more Items shall be present.  Each specimen imaged in the Pixel Data shall be identified by an Item in this Sequence. Other specimens in/on the container, but not imaged in the Pixel Data, may also be identified by Items in this Sequence.
...			
>Specimen Preparation Sequence	(0040,0610)	2	Sequence of Items identifying the process steps used to prepare the specimen for image acquisition. This includes description of all processing necessary to interpret the image. Zero or more Items shall be present.  This Sequence includes description of the specimen sampling step from an ancestor specimen, potentially back to the original part collection.  See Section C.7.6.2x.1.3.
>>Specimen Preparation Step Content Item Sequence	(0040,0612)	1	Sequence of Content Items identifying the processes used in one preparation step to prepare the specimen for image acquisition. One or more Items may be present.
>>>Include 'Content Item Macro' Table 10-2			Baseline Template ID is <del>81018001</del> Specimen Preparation
>Include 'Primary Anatomic Structure Macro' Table 10-8			Original anatomic location in patient of specimen. This location may be identical to that of the parent specimen, may be further refined by modifiers depending on the sampling procedure for this specimen, or may be a distinct concept.

>Specimen Localization Content Item Sequence	(0040,0620)	1C	Sequence of Content Items identifying the location of the specimen in the container and/or in the image. See Section C.7.6.2x.1.4.  Required if multiple specimens present in the image; may be present otherwise. One or more Items shall be present.
>>Include 'Content Item Macro' Table 10-2			Defined Template ID is <del>81048004</del> Specimen Localization